



NATIONAL HEALTH MISSION

Government of Meghalaya

No. DHS/MCH&FW/NHM/RCH/65/2021/3311

Dated Shillong the 22.08.22

NOTICE

Following Technical bid opening which was held on 06th/June/2022 pertaining to Tender Document for "Supply and Installation Of Blood Bank Equipments, Instruments and Others" vide Tender Notice No: DHS/MCH&FW/NHM/RCH/65/2021/3236, dated 02nd/May/2022, placed below the findings and decision thereafter:

Sl. no.	Bidder's Name	Remarks
1	M/S Arti Enterprise, Elke Drugs & Distributor, M/S R.K.S Enterprise, Areng Medical Supplier R.G.S. Enterprise	Bidders are qualified
2	Thermofisher	Did not furnish Medical equipments customer feedback a. Medical equipments customer feedback b. A valid Trade License Certificate from KHADC/JHADC/ GHADC for Non-Tribal firm c. TECHNICAL BID: Annexure – II a. Bidder is disqualified

Sample evaluation:

Item no.	Name of the Item	Name of Bidders				
		Arti Enterprise	R.K.S. Enterprise	Areng Medical Supplier	Elke Drugs Distributor	R.G.S. Enterprise
1	Deep Freezer -40°C	Complied	Complied	<p>Not complied:</p> <p>1. Krish Biomedicals: Temperature requirement upto 40°C the range of this equipment is 20°C to 25°C</p> <p>2. Acmas Technologies model ATI-127: Temperature chart recorder device not mentioned in catalogue. Detailed alarms 1-5 not mentioned</p> <p>3. Icon Instruments: Catalogue does not specify about</p> <p>1) Alarmer 1-5</p> <p>2) Temperature chart recorder</p> <p>4. Blue Star Limited: As per catalogue beep chart recorder facility not mentioned</p> <p>5. Diamond Scientific and Surgicals: Catalogue does not specify whether it is upreplit model & other features required for plasma storage</p> <p>6. Tempo Instruments Pvt. Ltd.: Temperature chart recorder details of & alarm not mentioned in catalogue</p> <p>Complied:</p> <p>1. Acmas Technologies Model No: Weiber Deep Freezer</p> <p>2. Middleby</p>	Complied	Upright model required. -Details of A/V alarms not provided.
2	Blood Bank Refrigerator	Complied	Labworld Scientific Instruments: Details of temp chart recorder or alarms not mentioned in catalogue, other company are complied	<p>Not complied:</p> <p>1. Acmas Technologies Model No: ATI-200: Temperature chart recording facility not specified</p> <p>2. Krish Biomedicals: Not as required as temperature range 3-16°C</p> <p>3. Narang Medical Ltd: Alarms 1-5 not specified in catalogue</p> <p>4. Diamond Scientific: Details of alarm not provided in catalogue</p> <p>Complied:</p> <p>1. Acmas Technologies: Model No: Weiber Blood Bank refrigerator (ACM - 78095 S-I)</p> <p>2. Middleby CelfrostInnovation</p>	Complied	Should provide with details of Temperature chart recorder & A/V Alarms not mentioned

Draft #53 of File DHS/MCH&FW/NHM/RCH/65/2021 Approved by Mission Director on 18-Aug-2022 05:48 PM - Page 1

Office of Mission Director, National Health Mission

Directorate of Health Services, Health Complex, Upper New Colony, Laitumkrah, Shillong - 793003

Phone: (0364) 2504532 Email: nrhmmegh@gmail.com

www.nrhmmeghalaya.nic.in

f Nhm Meghalaya

@iecbccnhmmegh

IECBCC NHM Meghalaya



NATIONAL HEALTH MISSION

Government of Meghalaya

Draft #53 of File DHS/MCH&FW/NHM/RCH/65/2021 Approved by Mission Director on 18-Aug-2022 05:48 PM - Page 2

3	Deep Freezer -80°C	Complied	Complied	<p>Not complied:</p> <ol style="list-style-type: none"> Blue Star Limited: Catalogue does not mention temperature recording chart facility Krish Biomedicals: No Details of ink chart recorder in catalogue Diamond Scientific : As per catalogue model is horizontal & not appropriate plasma storage Icon Instruments Company: Temperature range inadequate for plasma storage. <p>Complied:</p> <ol style="list-style-type: none"> Acmas Technologies Model No: ilabot & Model No: ATJ-127 Middleby CelfrostInnovation 	Complied	Storage capacity of 24-26 units required
4	Platelet Incubator	Complied	Complied	<p>Not complied:</p> <ol style="list-style-type: none"> Narang Medical Ltd: Capacity not specified in catalogue in terms of units -Agitator not mentioned in catalogue -Alarms 1-5 not mentioned in catalogue Deluxe Scientific Model : BBPC 101 : capacity 48bags whereas request is 24-26 bags, detailed Alv Alarms 1-5 not specified Deluxe Scientific Model : BBPC 111 : Ink chart recorder not available Span Healthcare: Alarm & sensor power failure not mentioned in catalogue Diamond Scientific : Details of alarm 1-5 not provided in catalogue <p>Complied:</p> <ol style="list-style-type: none"> Bioline India Middleby CelfrostInnovation 	Storage capacity for 24-26 bags required not 54 units	Complied
	Platelet Agitator	Complied	Labworld Scientific Instruments: Required capacity is 24-26 unit not 36/48 bags, other company are complied		Complied	
5	Semi Automated Coagulometer	Complied	<ol style="list-style-type: none"> 1 Channel Tests fibrinogen & factors not mentioned Printer not mentioned 	Complied	Complied	Complied
6	Plasma Thwing Bath	Complied	Complied	<p>Not complied:</p> <ol style="list-style-type: none"> Diamond Scientific and Surgicals: Temperature range not mentioned in catalogue <p>Complied:</p> <ol style="list-style-type: none"> Bioline India Narang Medical Ltd Span Healthcare Private Ltd 	Complied	Complied
7	Dielectric Tube Sealer	Complied	Complied	Complied	Complied	Complied
8	Blood Collection Monitor	Complied	Complied	Complied	Complied	Complied
9	Pre Vacuum Autoclave	Not Quoted	Not Quoted	Complied	Complied	Complied
10	Manual Plasma Expresser	Complied	Complied	Complied	Complied	Complied
11	Refrigerated Water Bath (Cryobath)	Complied	York Scientific Industries Pvt Ltd: Capacity not mentioned in terms of blood bag thawing, other company are complied	<p>Not complied:</p> <ol style="list-style-type: none"> Tempo Instruments : Temperature range inadequate Diamond Scientific and Surgicals: Temperature range not in catalogue <p>Complied:</p> <ol style="list-style-type: none"> Acmas Technologies Narang Medical Ltd Span Healthcare Private Ltd Bioline India Acmas Technologies 	Complied	Capacity of 10-20 bags per run/1 cycle is required
12	Elisa Reader and Washer	Complied	Complied	Complied	Complied	Complied
	Elisa Reader and Washer	Complied	Complied	Complied	Complied	Complied
13	Table Top Centrifuge	Complied	Labworld Scientific Instruments & Cosmo Laboratory Equipments: 1. Rotor Head for 16 tubes to be procured at extra cost.	<p>Not complied:</p> <ol style="list-style-type: none"> Deluxe Scientific: Capacity = 8/15 tubes and Requirement =16/24 tubes <p>Complied:</p>	Complied	Complied

Office of Mission Director, National Health Mission

Directorate of Health Services, Health Complex, Upper New Colony, Laitumkhrah, Shillong - 793003

Phone: (0364) 2504532 Email: nrhmmegh@gmail.com

www.nrhmmehalaya.nic.in

f Nhm Meghalaya



@iecbccnhmmegh



IECBCC NHM Meghalaya